

EMS CO-OPERATIVE HOSPITAL SOCIETY Q1665

Pathanapuram P.O, Kollam

Phone: 0475 -2354455

www.emshospital.org, emshospitalptpm@gmail.com

Tender No: PDT -1/2018-19/- EMS

AMENDMENT NOTICE-TENDER FOR SUPPLY OF HOSPITAL EQUIPMENTS&MEDICAL FURNITURE

Sub: Tender no PDT -1/2018-19/- EMS released on 24/05/2018 amendments published –reg

The amendments made in technical specifications are given below

SL NO	ITEM	Existing Clause	Amended Clause
1	Anesthesia Workstation	6.Should have monitoring facility of airway pressure, tidalvolume, frequency, oxygen concentration and AGM with modular integration. 7.Should have touch screen display of at least 10 inches for set parameters and graphical display for measured parameters.	Clause 6 removed. Should have touch screen display of minimum 8 inches for set parameters and graphical display for measured parameters
2	Multiparamonitor	1. Should be able to monitor ECG, SPO2, NIBP and respiration rate.	Should be able to monitor ECG,SPO2(Massimo/other),NIBP and respiration rate.

24/05/2018

Sd/-
SECRETARY